



Certificate of Compliance

BD Biosciences - Discovery Labware
One Becton Circle
Durham NC 27712 US

Page: 1 of 1

Product Name : Dish 100x15mm nonTC Petri 500cs
Catalog Number : 351029 **Manufacture Date** : 2012/10/06
Batch Number : 2279504
Expiration Date : 2017/10/31

Quality System Compliance - BD Falcon™ products are manufactured under the ISO 9001:2008 Standard, and the current FDA Quality System Regulation in Chapter 1, Title 21, Section 820 of the Code of USA Federal Regulations.

Sterilization - Product labeled as sterile is irradiated and dosimetrically released upon U.S. Association for the Advancement of Medical Instrumentation (AAMI) recommended practices in effect at the time of validation. Sterilization records are reviewed and signed off by qualified personnel for product release. BD Biosciences - Discovery Labware products labeled sterile meet a minimum requirement of 10^{-3} SAL (Sterility Assurance Level).

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and signed off by qualified personnel for product release.

Cytotoxicity - Testing is conducted to qualify all material resins using USP and/or ISO 10993 standards for cytotoxicity and have been shown to be non-toxic.

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

This product met BD Biosciences - Discovery Labware stringent quality standards at the time of batch/lot release. Any test results reported on this certificate were obtained at the time of release.

Tara Rogers
Quality Manager