

Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 1060-500

Lot Number: 22912002

Product Description: PYREX® Beaker, Berzelius, Tall Form, with Spout, Single Scale, 500 mL

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:

Capacity - 500 mL ±25 mL at any mark (5% of full capacity)

TSE/BSE:

Materials of animal origin are not used in the manufacture of this product.

This product is designed to comply with the requirements of ASTM E960 (Specification for Laboratory Glass Beakers). This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:

Visual Attributes
Dimensional
Packaging
Capacity

Results:

Pass
Pass
Pass
Pass

Revision Date:

06 MAR 2013



Arthur D. Birdsall
Quality Assurance Manager

Rev No: 1