

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

# Certificate of Compliance

**Catalog Number:** 5641-1L  
**Lot Number:** 32112032  
**Product Description:** PYREX® 1L Economy Volumetric Flasks, Glass Standard Taper Stopper  
**Component Materials:** Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E-438 (Standard Specification for Glasses in Laboratory Apparatus)  
**Attributes:** Capacity: 1000mL  $\pm$ 0.60mL

This product is designed to comply with the requirements of ASTM E 288 (Specification for Volumetric Flasks). This product is calibrated in accordance with ASTM E 542 (Specification for Calibration of Laboratory Volumetric Apparatus) to assure conformance to capacity specification. This product complies with all criteria for class B volumetric ware per ASTM E 694 (Specification for Volumetric Ware). This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following Characteristics:

<b>Test/Procedure:</b>	<b>Result:</b>
Visual Attributes	Pass
Dimensional Specification	Pass
Packaging	Pass
Capacity	Pass



Arthur D. Birdsall  
Quality Assurance Manager

Rev No: 82235