

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 5360-25

Lot Number: 21611220

Product Description: PYREX® Flask, Filtering, Micro, Tubulation, 25 mL
Tooled for Rubber Stopper #3

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:

Capacity - 25 mL

TSE/BSE:

Materials of animal origin are not used in the manufacture of this product.

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:

Visual Attributes
Dimensional
Packaging

Results:

Pass
Pass
Pass

Revision Date:

19 JUL 2013



Frigyes Szendro Kassa
Quality Manager

Rev No: 1