



Quality Certificate

VWR* GLOVES

This document certifies that the below product meets the following criteria:

| | |
|----------------------------|------------------------------|
| Product Number | 82026-424 |
| Lot Number | PT234R13C002 |
| Description | VWR GLOVE NITRILE PF S PK100 |
| Manufacturing Date: | March 2013 |
| Product Shelf Life: | February 2018 |

Material: Nitrile Butadiene Rubber , 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable product specification.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)
Meet or exceed EN 455-1&2
ISO 2859 (sampling table)

Accuracy:

| | |
|--------------------|---|
| Water tight test: | Inspection level=G1;AQL 1.5 |
| Visual inspection: | Inspection Level=G1;AQL 1.5 (criticaldefects) AQL 2.5 (major defects) AQL 4.0 (minor defects) |
| Dimensional: | Inspection Level= S2;AQL 4.0 |
| Physical property: | Inspection Level=S2; AQL 4.0 |
| Powder Content: | Inspection Level=5pieces ; Average \leq 2mg/glove |
| Moisture Content: | Inspection level=5pieces; Average \leq 1% |
| Glove Content: | Inspection level=5dispensers; Tolerance 100+/-2 pieces |

Sterility: N/A

Non-Pyrogenic: N/A

Rnase/Dnase Free: N/A

Regulatory: FDA 510K,ISO 9001,ISO 13485 under CMDCAS Quality System, GMP certificate

Signed:

Ken Crossley
Manager
Quality Assurance

Date : March 31, 2013