| Corning Incorporated Life Sciences | | |
|---------------------------------------|-----------|--|
| <u>Registered</u> ISO 9001:2008 | | Certificate of Compliance |
| Catalog Number: | 3025-25 | |
| Lot Number: | 21213130 | |
| Product Description: | PYREX® Cy | linder, Double Metric Scale, Economy, Graduated, TC, 25 mL |

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:

Capacity - 25 mL ±1.3 mL at any mark (5% of full capacity)

TSE/BSE:

Materials of animal origin are not used in the manufacture of this product.

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure: Visual Attributes Dimensional Packaging Results: Pass Pass Pass

Revision Date: 08 JUL 2013

Frigyes Szendro Kassa Quality Manager

Rev No: 1