

**Corning Incorporated  
Life Sciences**

**Registered**  
ISO 9001:2008

# Certificate of Compliance

**Catalog Number:** 4060-1L

**Lot Number:** 03114001

**Product Description:** PYREX® Flask, Florence, Boiling, Flat Bottom, Long Neck, 1000 mL  
Tooled for Rubber Stopper #8

**Lot Number Designation:**

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Component Materials:**

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

**Attributes:**

Capacity - 1000 mL

**TSE/BSE:**

Materials of animal origin are not used in the manufacture of this product.

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

**Test/Procedure:**

Visual Attributes  
Dimensional  
Packaging

**Results:**

Pass  
Pass  
Pass

**Revision Date:**

08 MAR 2013



Frigyes Szendro Kassa  
Quality Manager

Rev No: 1