



## **Quality Certificate**

## **VWR \* NITRILE EXAMINATION GLOVE**

This document certifies that the below product meets the following criteria:

North American Catalog No:	82026-428	
Lot No:	PT234R13S006	
Description:	VWR GLOVE NITRILE PF L PK100	
Manufacturing Date: December 2013		
Product Shelf Life :	November 2018	

Material:	Nitrile BUTADIENE Rubber, 100% Latex Free		
QC Testing:	<ul> <li>Representative production samples are collected and inspected in accordance with current applicable Product specification:</li> <li>Meet or exceed ASTM D 6319-00a (Water test &amp; Dimension)</li> <li>Meet or exceed EN 455 -1&amp;2</li> <li>ISO 2859 (Sampling table)</li> </ul>		
Accuracy:	Water tight test: Visual inspection:	Inspection level = G1; Inspection level = G1;	AQL 1.5 AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
	Dimensional: Physical Property:	Inspection level = S2; Inspection level = S2;	AQL 4.0 AQL 4.0
	Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
	Moisture Content: Glove count:	Inspection level = 5 pieces; Inspection level = 5 dispensers;	Average ≤ 1% Tolerance 100+/-2 pieces
Sterility	N/A		
Non- Pyrogenic:	N/A		
Rnase/ Dnase Free:	N/A		
Regulatory:	FDA 510K,ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate		

Signed:

Ken Crossley Manager Quality Assurance

Date: January 18, 2013