Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 5340-1L

Lot Number: 16214059

Product Description: PYREX® Flask, Filtering, Heavy Wall, Graduated, Tubulated, 1000 mL Tooled for Rubber Stopper #8

Lot Number Designation:
This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:
Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:
Capacity - 1000 mL ±50 mL at any mark (5% of full capacity)

TSE/BSE:
Materials of animal origin are not used in the manufacture of this product.

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Dimensional	Pass
Packaging	Pass

Revision Date:
19 JUL 2013



Racelis Acosta Reynoso
Plant Quality Manager

Rev No: 1