

## Quality Certificate

### VWR \* NITRILE EXAMINATION GLOVE

This document certifies that the below product meets the following criteria:

<b>North American Catalog No:</b>	82026-423
<b>Lot No:</b>	PT234R14E001
<b>Description:</b>	VWR GLOVE NITRILE PF XS PK100
<b>Manufacturing Date:</b>	May 2014
<b>Product Shelf Life :</b>	April 2019

**Material:** Nitrile BUTADIENE Rubber, 100% Latex Free

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

<b>Accuracy:</b>	Water tight test:	Inspection level = G1;	AQL 1.5
	Visual inspection:	Inspection level = G1;	AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
	Dimensional:	Inspection level = S2;	AQL 4.0
	Physical Property:	Inspection level = S2;	AQL 4.0
	Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
	Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
	Glove count:	Inspection level = 5 dispensers;	Tolerance 100+/-2 pieces

**Sterility** N/A

**Non- Pyrogenic:** N/A

**Rnase/ Dnase Free:** N/A

**Regulatory:** FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

**Signed:**



Ken Crossley  
Manager  
Quality Assurance

**Date:** May 13, 2014