



## **VWR \* NITRILE EXAMINATION GLOVE**

This document certifies that the below product meets the following criteria:

North American Catalog No:	82026-424
Lot No:	PT234R14J002
Description:	VWR GLOVE NITRILE PF S PK100
Manufacturing Date:	July 2014
Product Shelf Life :	June 2019

Material: Nitrile BUTADIENE Rubber, 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current

applicable Product specification:

Meet or exceed ASTM D 6319-00a (Water test & Dimension)

Meet or exceed EN 455 -1&2ISO 2859 (Sampling table)

Accuracy: Water tight test: Inspection level = G1; AQL 1.5

Visual inspection: Inspection level = G1; AQL 1.5 (critical defects) AQL 2.5 (major defects)

AQL 2.5 (major defects)

Dimensional: Inspection level = S2; AQL 4.0
Physical Property: Inspection level = S2; AQL 4.0

Powder Content: Inspection level = 5 pieces; Average ≤ 2 mg/glove

Moisture Content: Inspection level = 5 pieces; Average ≤ 1%

Glove count: Inspection level = 5 dispensers; Tolerance 100+/-2 pieces

Sterility N/A
Non- Pyrogenic: N/A
Rnase/ Dnase Free: N/A

Regulatory: FDA 510K,ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

Signed:

Ken Crossley Manager

Quality Assurance Date: July 21, 2014