

Corning Incorporated - Life Sciences  
1 Becton Circle  
Durham NC 27712 USA  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information.

Page: 1 / 1

<b>Product Name</b>	: Dish 35x10mm NonTC PetriEZGrip	<b>Manufacture Date</b>	: 2014-10-11
<b>Catalog Number</b>	: 351008		
<b>Batch Number</b>	: 4292019		
<b>Expiration Date</b>	: 2019-10-01		

**Quality System Compliance** - Falcon® products are manufactured under the ISO 9001: 2008 Standard, and the current FDA Quality System Regulation in Chapter 1, Title 21, Section 820 of the Code of USA Federal Regulations.

**Sterilization** - Product labeled as sterile is irradiated and dosimetrically released upon U.S. Association for the Advancement of Medical Instrumentation (AAMI) recommended practices in effect at the time of validation. Sterilization records are reviewed and signed off by qualified personnel for product release. Falcon products labeled sterile meet a minimum requirement of  $10^{-3}$  SAL (Sterility Assurance Level).

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and signed off by qualified personnel for product release.

**Cytotoxicity** - Testing is conducted to qualify all material resins using USP and/or ISO 10993 standards for cytotoxicity and have been shown to be non-toxic.

- The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

- This product met Corning Incorporated - Life Sciences stringent quality standards at the time of batch/lot release. Any test results reported on this certificate were obtained at the time of release.



Tara Rogers  
Quality Manager