

## VWR® TEST TUBE RACK

As per the manufacturer, the below product meets the following criteria:

<b>VWR Catalog Number</b>	89215-762
<b>Lot Number</b>	V301013
<b>Description</b>	VWR RACK 13MM-72-PL NYLON WHT
<b>Date of Manufacture:</b>	October 30, 2013
<b>Country of Origin:</b>	India.

### Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2008 & ISO13485:2003 Quality Management System. Inspection records are reviewed and signed for product release.

### QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

Test Name	Test Reference	Accuracy	Result
Autoclaving	TPPL/TEST/001	General Insp. Level- I, AQL- 1.5	PASS
Dimensional	QSF-QC-02	General Insp. Level- I, AQL- 1.5	Within Tolerance

### Product Specifications

**Material:** 30 % Glass filled Nylon with color masterbatch.

**BSE/TSE Statement:** The above mentioned product is manufactured without the use of Raw materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's knowledge.

**Non-Pyrogenic:** Not Applicable.

**DNase & RNase Free:** Not Applicable.

**Cytotoxicity:** Not Applicable.

**Latex Free Statement:** No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.

**Signed:**



Ken Crossley  
Manager  
Quality Assurance

**Date: October 30, 2013**