

Quality Certificate

VWR® NITRILE EXAMINATION GLOVE

As per the manufacturer, the below product meets the following criteria:

| | |
|------------------------------------|---|
| North American Catalog No : | 82026-428 |
| Lot No : | PT15P306 |
| Description : | VWR GLOVE NITRILE PF L PK100 |
| Date of Manufacturing : | November 2015 |
| Product Shelf Life : | October 2020 |
| Country of Origin : | Indonesia |
| Quality System Compliance : | FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate |

QC Testing:

Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

Accuracy:

| | | |
|--------------------|----------------------------------|----------------------------|
| Water tight test: | Inspection level = G1; | AQL 1.5 |
| Visual inspection: | Inspection level = G1; | AQL 1.5 (critical defects) |
| | | AQL 2.5 (major defects) |
| | | AQL 4.0 (minor defects) |
| Dimensional: | Inspection level = S2; | AQL 4.0 |
| Physical Property: | Inspection level = S2; | AQL 4.0 |
| Powder Content: | Inspection level = 5 pieces; | Average \leq 2 mg/glove |
| Moisture Content: | Inspection level = 5 pieces; | Average \leq 1% |
| Glove count: | Inspection level = 5 dispensers; | Tolerance 100+/-2 pieces |

Product Specifications:

| | |
|----------------------------|---|
| Material: | Nitrile Butadiene Rubber, Not made with natural rubber latex. |
| Sterility | N/A |
| Non- Pyrogenic: | N/A |
| RNase/ DNase Free: | N/A |
| Cytotoxicity: | N/A |
| BSE/ TSE Statement: | N/A |

Signed :



Ken Crossley
Manager
Quality Assurance

Date : 23 November 2015