



Becton Dickinson and Company BD Diagnostic Systems PO Box 999 Sparks MD 21152-0999 US

Page: 1 of 3

Product Name : Bottle Mannitol Salt Agar BBL 500G Catalog Number : 211407 Manufacture Date :2016/10/04 Batch Number : 6357520 **Expiration Date** : 2020/09/30 01. Dehydrated Medium Powder Appearance: Fine, homogeneous, free of extraneous material and may contain many light to dark red flecks 02. Solubility (111q/L): Soluble in distilled or deionized water on boiling. 03. Prepared Plate Appearance: Medium to dark, red to rose; clear to slightly hazy 04. USP/JP/EP performance specifications where applicable CULTURE ATCC® INOC RESULT INCUBATION TEMP30-35°C 18 Hrs Staphylococcus aureus 6538 <100 CFUs growth* yellow zones* 30-35°C 48-72 Hrs 8739 >100 CFUs no growth* 30-35°C 72 Hrs Escherichia coli * Growth results above are comparable to a previouly released batch. 05. Biological Performance: Medium was prepared per label directions. Plates were inoculated with the test organisms, incubated at 33-37°C for 42-48 hours and gave cultural responses as indicated: CULTURE ATCC® RESULT Proteus mirabilis 12453 Partial to complete inhibition Staphylococcus aureus 25923 Light to heavy growth, yellow zones at 48 hours Staphylococcus epidermidis 12228 Light to heavy growth, red zones at 48 hours

Characteristic	Unit	Value	Lower Limit	Upper Limit
pH at 25°C : Bulk Lot Number :	_	7.4 6252834	7.2	7.6

Animal source	Country of	Tissue Category		
	Origin	BIC	SIC	ABC
Bovine	New Zealand	IV	IV	MLK
Bovine	USA	II	II	IB
Bovine	USA	III	III	IB
Bovine	USA	III	IV	IC
Bovine	USA	IV	IV	IB
Equine	Mexico	II	II	IB
Equine	Mexico	IV	IV	IB
Porcine	Canada	IV	IV	IB
Porcine	USA	II	II	IB
Porcine	USA	III	III	IB
Porcine	USA	IV	IV	IB



Becton Dickinson and Company BD Diagnostic Systems PO Box 999 Sparks MD 21152-0999 US

Page: 2 of 3

Product Name	: Bottle Mannitol Salt Agar BBL 500G	
Catalog Number	: 211407 Manufacture Date :2016/10/0)4
Batch Number	: 6357520	
Expiration Date	: 2020/09/30	

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostics - Diagnostic Systems products are manufactured in ISO 9001:2008 Registered facilities. In addition, BD Diagnostics -Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), are regulated by the FDA's Quality System Regulations (QSRs), and are also ISO 13485:2003 Registered. This product met BDDS stringent quality standards at time of batch/lot Any test results reported on this certificate were obtained release. at time of release. This material is not for human or animal consumption. BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at http://www.bd.com/regdocs. For each batch of finished product that contains animal origin raw materials, the COA shows the animal origin data from the individual lots of animal origin raw materials used, as provided by the raw material suppliers. At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at http://www.bd.com/regdocs under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program. For complete details on animal origin information, refer to "BD Position Statement - BD Diagnostic-Diagnostic Systems, COA Animal Origin Information Position Statement", at http://www.bd.com/regdocs under "Position Statements". Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see www.bd.com/ds/technicalCenter/regulatory.asp. Stability data for this product is on file. The pH stated was obtained at the time of release, and the pH may vary within the stated range depending on the age of the product and the type of pH meter and probe used.



Becton Dickinson and Company BD Diagnostic Systems PO Box 999 Sparks MD 21152-0999 US

Page: 3 of 3

Product Name: Bottle Mannitol Salt Agar BBL 500GCatalog Number: 211407Manufacture Date :2016/10/04Batch Number: 6357520Expiration Date: 2020/09/30

Charlette Danserfilter

Charlotte Dannenfelser BD Life Sciences - Diagnostic Systems Quality Director, Microbiology Signature Date:2017/02/03