



661 Route 23 South, Wayne, New Jersey 07470

Tel: 973-694-0500 * Fax: 973-694-7199

Internet Address: <http://www.bel-art.com>

CERTIFICATE OF ANALYSIS

This is to verify that all items noted below are in conformance with the Purchase Order Contract, drawings, specifications, and other applicable documentation; that all process certifications, material certifications, are on file at this or another Bel-Art facility and are available for review by the customer and or Government Representatives upon request.

Product #: 28461-2000 **Mfg. Date:** 2/5/2017 **Lot #:** J291328

Description: Holdfast Graduated Cylinders 2000ml

Material: Cylinder: Polypropylene, FDA
Handle: Polypropylene, 20% Talc White, FDA

Dimensions: Overall height of 19.31", base dia @ max 6.35", dia. 3.43" OD
3.19" ID. Inside ht 18.72" Scale is in 20ml subdivisions.
Handles have a hand opening of 3" x 1 1/8".

Other: Graduated cylinders have a total contained, total delivered scale on two
sides, one ascending and one descending, with raised numerals. Large spouts
facilitate pouring and hexagon base provides stability. Non wetting surface
eliminates concave meniscus and gives more accurate readings.
Chemical resistant and steam autoclavable at 121° C. Class B, +/- 2ml accuracy
Based on the process and materials involved in the manufacturing of this
product, this product is considered to be BSE/TSE protein free
Manufacture in Pocomoke City, MD USA

Reviewed and Approved By:

Signature: Michael E. Grzelak

Printed Name: Michael E. Grzelak

Title: Quality Assistant

Date: 3/3/2017

SUBSIDIARIES: MADDAK, INC * APPLIED COATING, INC
ADDITIONAL PLANTS IN POCOMOKE, MD * KANSAS CITY, MISSOURI