

Date of Certificate: 05/05/2017
Product Number: 5000-0020
Product Lot Number: 1200148
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1943449	04/24/2017	156	CS	20.3	26.7
1948001	05/02/2017	156	CS	20.3	26.2



Barry Srolis
Sr. Quality Manager