

Corning Incorporated - Life Sciences  
 Antigua Carretera A Roma KM 3.6  
 San Nicolas De Los Garza NL 66480 MEX  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
 Refer to website for regional contact information.

Page: 1 / 1

<b>Product Name</b>	: 5580 FLASK 100ML	<b>Manufacture Date</b>	: 2017-04-03
<b>Catalog Number</b>	: 5580-100		
<b>Lot ID</b>	: 09317007		
<b>Expiration Date</b>	: Unconfirmed Validity		

**Quality Management System** - Complies with the current version of the ISO 9001 Standard.

**Glass Type** - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

**Volumetric Tolerance** - Volumetric Accuracy at fill line +/- 0.08 mL.

**Product Specification Standards** - Product made in accordance with:

ASTM E288 - Standard Specification for Laboratory Glass Volumetric Flasks

ASTM E542 - Specification for Calibration of Laboratory Volumetric Apparatus

ASTM E694 - Standard Specification for Laboratory Glass Volumetric Apparatus, Class A  
 Criteria

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass

Packaging Inspection - Pass

Dimensional Test - Pass

Capacity Verification - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Miguel Flores  
 Quality Manager