
ThermoFisher SCIENTIFIC

Certificate of Processing

Date of Certificate: 06/23/2017
Product Number: 5000-0020
Product Lot Number: 1202683
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1968404	05/31/2017	6	CS	19.9	25.9
1968417	06/05/2017	306	CS	20.3	26.7



Barry Srolis
Sr. Quality Manager

75 Panorama Creek Drive Rochester, NY 14625 Tel:585.586.8800 Fax:585.899.7605 www.thermofisher.com