

Legal Manufacturer:  
Becton Dickinson and Company  
1 Becton Drive  
Franklin Lakes, NJ 07417  
USA

Manufacturing Site:  
BD Medical - Diabetes Care  
1329 West Hwy 6  
Holdrege, NE 68949-0860  
USA



## **CERTIFICATE OF COMPLIANCE**

### **Issued by Manufacturing Site**

This is to certify that this lot of non-sterile Alcohol Swab: Lot Number: 6160532

Date of manufacture: 01 SEP 2016 Expiration Date: 31 AUG 2021

Catalog Number: 326895

Containing Isopropyl Alcohol Lot/s: 23160-BULK-22086P

<u>DOCUMENT NUMBER</u>	<u>REVISION</u>	<u>DOCUMENT DESCRIPTION</u>
SP700008	35	B-D Disposable Alcohol Swab (Master Production Document)
MD700533	35	Alcohol Swab Assembly and Packaging
QC700960	52	Process Inspection & Stat Sample – Single Use Swabs

Using the above documents and procedures, this product was produced under the U.S. Food and Drug Administration (FDA) current Good Manufacturing Practice (cGMP) at BD Medical - Diabetes Care, 1329 West Highway 6, Holdrege, Nebraska, 68949.

Sara Dubrow [Signature] 20 Sep 2016  
QUALITY ENGINEER – Printed name QUALITY ENGINEER – Signature DATE

## **CERTIFICATE OF ANALYSIS**

### **Issued By Manufacturing Site**

<u>TEST</u>	<u>TEST METHOD</u>	<u>SPECIFICATION</u>	<u>RESULTS</u>
Specific Gravity	HMIC-43	0.872 – 0.883	<u>0.8768g</u>
Assay	HMIC-43	68.0 - 72.0 %	<u>70.1%</u>

MICHAEL PUTERA [Signature] 09 Sep 2016  
LABORATORY – QUALITY ASSURANCE SUPERVISOR LABORATORY – QUALITY ASSURANCE SUPERVISOR DATE  
Printed Name Signature