

Legal Manufacturer:
Becton Dickinson and Company
1 Becton Drive
Franklin Lakes, NJ 07417
USA

Manufacturing Site:
BD Medical - Diabetes Care
1329 West Hwy 6
Holdrege, NE 68949-0860
USA



CERTIFICATE OF COMPLIANCE **Issued by Manufacturing Site**

This is to certify that this lot of non-sterile Alcohol Swab: Lot Number: 7214559

Date of manufacture: 01 OCT 2017 Expiration Date: 30 SEP 2022

Catalog Number: 326895

Containing Isopropyl Alcohol Lot/s: 23160-BULK-19107P/23160-BULK-27107P

<u>DOCUMENT NUMBER</u>	<u>REVISION</u>	<u>DOCUMENT DESCRIPTION</u>
SP700008	36	B-D Disposable Alcohol Swab (Master Production Document)
MD700533	35	Alcohol Swab Assembly and Packaging
QC700960	54	Process Inspection & Stat Sample – Single Use Swabs

Using the above documents and procedures, this product was produced under the U.S. Food and Drug Administration (FDA) current Good Manufacturing Practice (cGMP) at BD Medical - Diabetes Care, 1329 West Highway 6, Holdrege, Nebraska, 68949.

<u>Jami Holloway</u>	<u>Jami Holloway</u>	<u>05 Dec 2017</u>
QUALITY ENGINEER – Printed name	QUALITY ENGINEER – Signature	DATE

CERTIFICATE OF ANALYSIS **Issued By Manufacturing Site**

<u>TEST</u>	<u>TEST METHOD</u>	<u>SPECIFICATION</u>	<u>RESULTS</u>
Specific Gravity	HMIC-43	0.872 – 0.883	<u>0.8757g</u>
Assay	HMIC-43	68.0 - 72.0 %	<u>70.2%</u>

<u>MICHAEL PETERA</u>	<u>[Signature]</u>	<u>01 Dec 2017</u>
LABORATORY – QUALITY ASSURANCE SUPERVISOR Printed Name	LABORATORY – QUALITY ASSURANCE SUPERVISOR Signature	DATE