

Certificate of Analysis

Page: 1 of 3

Becton Dickinson and Company BD Diagnostic Systems PO Box 999 Sparks MD 21152-0999 US

Product Name : Bottle Tryptic Soy Agar 500G

Catalog Number : 236950 Manufacture Date: 2018/01/05

Batch Number : 8032542
Expiration Date : 2022/12/31

- 01. Dehydrated Medium Appearance: Light beige, free-flowing, homogeneous
- 02. Solubility: 4.0% solution, soluble in distilled or deionized water on boiling
- 03. Solution Appearance: Light amber, slightly to moderately opalescent
- 04. Plate Appearance: Clear to slightly opalescent, without precipitate
- 05. CAMP Test: Streptococcus agalactiae ATCC® 12386 gave a + reaction and Streptococcus pyogenes ATCC® 19615 gave a reaction when tested with Staphylococcus aureus ATCC® 33862.
- 06. USP/EP/JP Growth Promotion test performance specification where applicable.

TEST ORGANISMS	ATCC®	II	10C	RECOVER'	Y TEMP	11	ICUE	A	CION
Aspergillus	16404	<100	CFUs	growth	30-35°C	Uр	to	5	days
brasiliensis									
Bacillus subtilis	6633	<100	CFUs	growth	30-35°C	Uр	to	3	days
Candida albicans				growth	30-35°C	Uр	to	5	days
Escherichia coli	8739	<100	CFUs	growth	30-35°C	Uр	to	3	days
Pseudomonas aeruginosa				growth	30-35°C	Uр	to	3	days
Salmonella typhimurium	14028	<100	CFUs	growth	30-35°C	Uр	to	3	days
Staphylococcus aureus	6538	<100	CFUs	growth	30-35°C	Up	to	3	days

07. Cultural Response: Medium was prepared per label instructions. Plates with and without 5% sheep blood (SB) were inoculated with the test organisms and incubated under 5-10% CO2 at 35 ± 2°C for 18-48 hours.

io io ilouib:				
TEST ORGANISMS	ATCC®	RECO	VERY	HEMOLYSIS
		PLAIN	w/SB	
Escherichia coli	25922	good	good	beta
Neisseria meningitidis	13090	good	good	none
Staphylococcus aureus	25923	good	good	beta
Streptococcus pneumoniae	6305	good	good	alpha
Streptococcus pyogenes	19615	good	good	beta

Characteristic	Unit	Value	Lower Limit	Upper Limit
pH at 25°C : Bulk Lot Number :	_	7.3 8003886	7.1	7.5

Animal source	Country of	Tissue Category		
	Origin	BIC	sic	ABC
Bovine	New Zealand	IV	IV	MLK

Creation Date: 2018/02/14 16:26:00



Certificate of Analysis

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999

Sparks MD 21152-0999 US Page: 2 of 3

Product Name : Bottle Tryptic Soy Agar 500G

Catalog Number : 236950 Manufacture Date: 2018/01/05

Batch Number : 8032542 **Expiration Date** : 2022/12/31

Porcine USA III III IB

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostics - Diagnostic Systems products are manufactured in ISO 9001:2008 Registered facilities. In addition, BD Diagnostics - Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), are regulated by the FDA's Quality System Regulations (QSRs), and are also ISO 13485:2003 Registered. This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at http://www.bd.com/regdocs. For each batch of finished product that contains animal origin raw materials, the COA shows the animal origin data from the individual lots of animal origin raw materials used, as provided by the raw material suppliers.

At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at http://www.bd.com/regdocs under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program.

For complete details on animal origin information, refer to "BD Position Statement - BD Diagnostic-Diagnostic Systems, COA Animal Origin Information Position Statement", at http://www.bd.com/regdocs under "Position Statements".

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see www.bd.com/ds/technicalCenter/regulatory.asp.

Creation Date: 2018/02/14 16:26:00



Certificate of Analysis

Becton Dickinson and Company BD Diagnostic Systems PO Box 999

Sparks MD 21152-0999 US Page: 3 of 3

Product Name : Bottle Tryptic Soy Agar 500G

Catalog Number : 236950 Manufacture Date: 2018/01/05

Batch Number : 8032542 **Expiration Date** : 2022/12/31

Charlette Danserfler

Charlotte Dannenfelser
BD Life Sciences - Diagnostic Systems
Quality Director, Microbiology
Signature Date: 2018/02/14

Creation Date: 2018/02/14 16:26:00