



## **VWR® POWDER-FREE NITRILE EXAMINATION GLOVES**

As per the manufacturer, the below product meets the following criteria:

North American Catalog No :	82026-426
Lot No :	PT18D205
Description :	VWR GLOVE NITRILE PF M PK100
Date of Manufacturing :	April 2018
Product Shelf Life :	March 2023
Country of Origin :	Indonesia

## **Quality System Compliance:**

Products are manufactured under FDA510K, ISO 9001:2008, ISO 13485:2003I, CMDCAS Quality System standard and Good Manufacturing Practices. Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

<u>QC Testing:</u>	Representative production samples are collected and inspected in accordance with current applicable Product specification:			
<ul> <li>Meet or exceed ASTM D6319 and ISO 2859 (Sampling table)</li> </ul>				
<u>Accuracy:</u>	Water tight test: Visual inspection:	AQL	1.5 1.5 (critical defects) 2.5 (major defects) 4.0 (minor defects)	
	Dimensional: Physical Property Powder Content: Moisture Content: Glove count:	Inspection level = 5 pieces;AverageInspection level = 5 pieces;Average		
Product Specification:		This product is Class I category device as defined by the FDA in 21CFR parts 807 sub parts C		
Material:		Nitrile Butadiene Rubber.		
Non- Pyrogenic:		N/A		
Dnase & RNase Free Statement :		N/A		
Cytotoxicity Statement:		N/A		
BSE/ TSE Statement:		N/A		
Latex Statement:		This product is not made with natural rubber latex		

This document has been produced electronically and is valid without a signature.

## Signed on behalf of VWR International:

Michael J. Mascali, RPh, FASCP, ASQ CMQ – OE SVP Compliance VWR International LLC

Date : May 05, 2018