

Certificate Of Analysis

Whatman™

Product Information

Product Number: 28320-085

Product Name: 415 11CM 100/PK

Lot Number/Serial Number: 16893578

Date of Manufacture: 20-Jul-2018

Quality Control

Test	Range	Results
Grammage,gsm	68-84	80
Thickness, um/53kpa	195-325	287
Porosity, S/300ml/0.1sq.in	40-80	70

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to ISO9001:2015 and tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

Electronic signature

This document has electronically produced and is valid without a signature.

Version AC



www.gelifesciences.com

GE Healthcare UK Limited

Amersham Place Little Chalfont
Buckinghamshire HP7 9NA UK