

VWR® LABORATORY BOTTLES, POLYPROPYLENE, NARROW MOUTH

As per the manufacturer, the below product meets the following criteria:

| | |
|-----------------------------|----------------------------|
| VWR Catalog Number | 414004-178 |
| EU Article Number | 215-2290 |
| Lot Number | VAF191116 |
| Description | VWR BOTTLE PP NM 30ML PK12 |
| Date of Manufacture: | November 19, 2016. |
| Country of Origin | India. |

Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2008 & ISO13485:2003 Quality Management System. Inspection records are reviewed and signed for product release.

QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

| Test Name | Test Reference | Accuracy | Result |
|--------------------|----------------|----------------------------------|------------------|
| Leak Test | QWI-QC-03 | General Insp. Level- I, AQL- 1.5 | Leak proof |
| Autoclaving | QWI-QC-07 | General Insp. Level- I, AQL- 1.5 | Autoclavable |
| Dimensional | QSF-QC-02A | General Insp. Level- I, AQL- 1.5 | Within tolerance |

Product Specifications

Material: **Bottle:** Polypropylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.

Closure: Polypropylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.

BSE/TSE Statement: The above mentioned product is manufactured without the use of Raw materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's knowledge.

Non-Pyrogenic: Not Applicable.

DNase & RNase Free: Not Applicable.

Cytotoxicity: Not Applicable.

Latex Free Statement: No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.

Signed:



Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: October 17, 2018.