

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 3413

Lot Number: 02813041

Product Description: Costar® Transwell®, 6.5mm, TC treated, 0.4 µm pore size, in a 24-well cluster plate with lid

Component Materials:

- | | | |
|----------------|---|--|
| Plate/Lid | - | Virgin Polystyrene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |
| Transwell body | - | Virgin Polystyrene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |
| Filter | - | Translucent polycarbonate membrane, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10⁻³

Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of Cell Attachment and Growth utilizing an attachment- dependent mammalian cell line in a serum supplemented media.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:


Test/Procedure:	Results:
Visual Attributes	Pass
Cell Culture Treatment	Pass
Packaging	Pass

Expiration Statement:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.



Judith A. Gibbs
Quality Assurance Manager

Rev No: 4