



Quality Certificate

VWR* EXAM GLOVES

This document certifies that the below product meets the following criteria:

North American Catalog No:	82026-426
Lot Number:	234R10J008
Description:	VWR GLOVE NITRILE PF M PK100
Manufacturing date:	8.3.2010
Expiration date(when applicable):	8.3.2010

Material : SF56g/m2 breathable fabric.

QC Testing : Representative production samples are collected and inspected in accordance with current applicable product specifications:

Break Tensile properties(N): ASTM D5034 standard, MD=101.19 CD=100.28

Elongation rate (%) : ASTM D5034 standard, MD=114.99 CD=133.05

Trapezoid tearing strength(N): FZ/T01085-1999 standard, MD=556 CD=749

Irradiated Date (when applicable): 2010-7-18

Irradiation lot number(when applicable): 1007020008

Sterilization (when applicable) : Gamma-ray irradiated according to ISO11137:2006, Dose set meet SAL 10^{-6} , With the dose range of : Min. 17.0KGY Max. 22.9KGY

Latex free statement (when applicable) : The above product is latex free.

Origin : Made in China.

Signed:

Ken Crossley
Manager
Quality Assurance