

Certificate of Analysis

Becton Dickinson Caribe LTD. PO Box 372860 Cayey PR 00737-2860 PR

Page: 1 of 1

Product Name : DISC BACITRACIN B-10 10 EA

Catalog Number : 231268 Manufacture Date : 2011/08/03

Batch Number : 1199230
Expiration Date : 2014/08/31

| Characteristic | Unit | Value | LowLimit | HighLimit |
|--------------------|------|-------|----------|-----------|
| Micro Assay Range: | | 12.68 | | |
| Micro Assay Range: | | 12.54 | | |
| Micro Assay Range: | | 12.85 | | |
| Micro Assay Range: | | 12.42 | | |
| Micro Assay Potenc | | 13.00 | | |
| Uniformity | mm | 0.6 | - | 2.5 |
| Uniformity | mm | 1.0 | _ | 2.5 |
| Uniformity | mm | 0.6 | - | 2.5 |
| Uniformity | mm | 0.6 | - | 2.5 |

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostic Systems (BDDS) is an ISO 13485:2003 and ISO 9001:2008 Registered facility. BDDS products are manufactured in facilities registered with the United States Food and Drug Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release.

John Gerlich Vice President,

Quality Management and Regulatory Compliance

Signature Date: 2011/09/13

Creation Date: 2011/09/13 19:43:27