



Quality Certificate

VWR* GLOVES

This document certifies that the below mentioned product meets the following criteria:

| | |
|----------------------------|------------------------------|
| Product Number | 82026-428 |
| Lot Number | PT234R13E005 |
| Description | VWR GLOVE NITRILE PF L PK100 |
| Manufacturing Date: | May 2013 |
| Product Shelf Life: | April 2018 |

Material: Nitrile BUTADIENE Rubber , 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable product specification.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)
Meet or exceed EN 455-1&2
ISO 2859 (sampling table)

Accuracy:

| | |
|--------------------|-----------------------------------------------------------------------------------------------------|
| Water tight test: | Inspection level=G1; AQL 1.5 |
| Visual inspection: | Inspection Level=G1;AQL 1.5 (criticaldefects) AQL 2.5 (major defects) AQL 4.0 (minor defects) |
| Dimensional: | Inspection Level= S2; AQL 4.0 |
| Physical property: | Inspection Level=S2; AQL 4.0 |
| Powder Content: | Inspection Level=5pieces ; Average ≤ 2 mg/glove |
| Moisture Content: | Inspection level=5pieces; Average $\leq 1\%$ |
| Glove Content: | Inspection level=5 dispensers; Tolerance 100+/-2 pieces |

Sterility: N/A

Non-Pyrogenic: N/A

Rnase/Dnase Free: N/A

Regulatory: FDA 510K,ISO 9001,ISO 13485, CMDCAS Quality System, GMP certificate

Signed:

Ken Crossley
Manager
Quality Assurance

Date : March 22, 2013