Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 3022-25

Lot Number: 02012045

Product Description: PYREX® Cylinder, Single Metric Scale, Graduated, TC, 25 mL with LDPE Bumper Guard

Lot Number Designation:
This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:
Glass - Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)
Guard - Low Density Polyethylene (LDPE)

Attributes:
Capacity - 25 mL ±0.34 mL at any mark

This product is designed to comply with the requirements of ASTM E133 (Specification for Distillation Equipment), ASTM E1272 (Specification for Graduated Cylinders) and class B criteria per ASTM E694 (Specification for Volumetric Ware). It is calibrated in accordance with ASTM E542 (Specification for Calibration of Laboratory Volumetric Apparatus). This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Dimensional	Pass
Packaging	Pass
Capacity	Pass

Revision Date:
03 JUL 2013



Frigyes Szendro Kassa
Quality Manager

Rev No: 1