

Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 7077-5N
Lot Number: 24712746
Product Description: PYREX® Pipet, Disposable Glass, Serological, 5 mL in 1/10 mL
Individual Wrap, Sterile, Plugged

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:

Glass - Type I Class B - Alumino-Borosilicate Glass,
per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)
Plug - Polyester (13 grain)

Attributes:

Capacity - 5.0 mL $\pm$ 0.15 mL at any mark (3% of full capacity)

Sterilization:

Product sterilized by Ethylene Oxide (ETO) per ANSI/AAMI/ISO 11135-1994, ANSI/AAMI/ISO 11137-1994 and EN 550/552 requirements.

TSE/BSE:

Materials of animal origin are not used in the manufacture of this product.

This product is designed to comply with the requirements of ASTM E 714 (Specification for Disposable Glass Serological Pipets). This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:

Visual Attributes
Dimensional
Packaging
Capacity

Results:

Pass
Pass
Pass
Pass

Revision Date:

9 OCT 2013

Arthur D. Birdsall
Quality Assurance Manager

Rev No: 2