



Quality Certificate

VWR* GLOVES

This document certifies that the below product meets the following criteria:

Product Number	82026-423
Lot Number	PT234R14B001
Description	VWR GLOVE NITRILE PF XS PK100
Manufacturing Date:	March 2014
Product Shelf Life:	February 2019

Material: Nitrile Butadiene Rubber, 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable product specification.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)
Meet or exceed EN 455-1&2
ISO 2859 (sampling table)

Accuracy:

Water tight test:	Inspection level=G1; AQL 1.5
Visual inspection:	Inspection Level=G1; AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
Dimensional:	Inspection Level= S2; AQL 4.0
Physical property:	Inspection Level=S2; AQL 4.0
Powder Content:	Inspection Level=5pieces ; Average ≤2mg/glove
Moisture Content:	Inspection level=5pieces; Average≤1%
Glove Count:	Inspection level=5dispensers; Tolerance 100+/-2 pieces

Sterility: N/A

Non-Pyrogenic: N/A

Rnase/Dnase Free: N/A

Regulatory: FDA 510K, ISO 9001, ISO 13485 under CMDCAS Quality System, GMP certificate

Signed:

Ken Crossley
Manager
Quality Assurance

Date: March 04 , 2013