



Becton Dickinson and Company  
BD Diagnostic Systems  
PO Box 999  
Sparks MD 21152-0999 US

# Certificate of Analysis

Page: 1 of 3

**Product Name** : BOTTLE TRYPTIC SOY AGAR 500G  
**Catalog Number** : 236950 **Manufacture Date** : 2011/01/10  
**Batch Number** : 1026616  
**Expiration Date** : 2015/12/31

01. Dehydrated Medium Appearance: Light beige, free-flowing, homogeneous  
02. Solubility: 4.0% solution, soluble in distilled or deionized water on boiling  
03. Solution Appearance: Light amber, slightly opalescent  
04. CAMP Test: Streptococcus agalactiae ATCC® 12386 gave a + reaction and Streptococcus pyogenes ATCC® 19615 gave a - reaction when tested with Staphylococcus aureus ATCC® 33862.  
05. USP/EP/JP Growth Promotion test performance specification where applicable.

TEST ORGANISMS	ATCC®	INOC	RECOVERY	TEMP	INCUBATION
Aspergillus niger	16404	<100 CFUs	growth	30-35°C	Up to 5 days
Bacillus subtilis	6633	<100 CFUs	growth	30-35°C	Up to 3 days
Candida albicans	10231	<100 CFUs	growth	30-35°C	Up to 5 days
Escherichia coli	8739	<100 CFUs	growth	30-35°C	Up to 3 days
Pseudomonas aeruginosa	9027	<100 CFUs	growth	30-35°C	Up to 3 days
Salmonella typhimurium	14028	<100 CFUs	growth	30-35°C	Up to 3 days
Staphylococcus aureus	6538	<100 CFUs	growth	30-35°C	Up to 3 days

06. Cultural Response: Medium was prepared per label instructions. Plates with and without 5% sheep blood (SB) were inoculated with the test organisms and incubated under 5-10% CO<sub>2</sub> at 35 ± 2°C for 18-48 hours.

TEST ORGANISMS	ATCC®	RECOVERY	HEMOLYSIS
		PLAIN	w/SB
Escherichia coli	25922	good	good
Neisseria meningitidis	13090	good	good
Staphylococcus aureus	25923	good	good
Streptococcus pneumoniae	6305	good	good
Streptococcus pyogenes	19615	good	good

07. Residual Solvents (CPMP/ICH/283/95): Typical Analysis for Tryptic Soy Agar indicates that there is less than 5000 ppm of Acetone. No other solvents were detected during analysis.

Characteristic	Unit	Value	LowLimit	HighLimit
pH at 25°C :		7.3	7.1	7.5
Bulk Lot Number :	-	0351780		



Becton Dickinson and Company  
BD Diagnostic Systems  
PO Box 999  
Sparks MD 21152-0999 US

## Certificate of Analysis

Page: 2 of 3

**Product Name** : BOTTLE TRYPTIC SOY AGAR 500G  
**Catalog Number** : 236950 **Manufacture Date** : 2011/01/10  
**Batch Number** : 1026616  
**Expiration Date** : 2015/12/31

Animal Source	Country of Origin	Tissue Category		ABC
		BIC	SIC	
Porcine	USA	III	III	B
Bovine	New Zealand	IV	IV	C
Porcine	Canada	III	III	B

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostic Systems (BDDS) is an ISO 13485:2003 and ISO 9001:2008 Registered facility. BDDS products are manufactured in facilities registered with the United States Food and Drug Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release.

BD Diagnostic Systems' Certificates of Analysis (COA) typically are set up to contain animal origin information for finished products manufactured using materials of animal origin. The animal origin information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. This information is a compilation of animal origin data from the individual lots of raw materials used to manufacture the batch of BD Diagnostic Systems (BDDS) finished product for which the COA was created.

At the time the BDDS Certificate of Analysis is created and sent to the Internet website address at <http://www.bd.com/regdocs/>, the animal origin information as provided to BDDS by its suppliers is pulled into the certificate as it is created by the BDDS automated certificate system.

At times, suppliers notify BDDS of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BDDS. When this situation occurs, BDDS updates the animal origin information in the automated certificate system, recreates the affected finished product COAs for batches within expiration date, and sends them to the Internet website where they replace the prior certificate and are immediately available to customers.

Customers enrolled in BD Diagnostic Systems' Automated Change Notification Program will be notified of the changes described above.



Becton Dickinson and Company  
BD Diagnostic Systems  
PO Box 999  
Sparks MD 21152-0999 US

## Certificate of Analysis

Page: 3 of 3

**Product Name** : BOTTLE TRYPTIC SOY AGAR 500G  
**Catalog Number** : 236950 **Manufacture Date** : 2011/01/10  
**Batch Number** : 1026616  
**Expiration Date** : 2015/12/31

For complete details refer to "BD Position Statement - BD Diagnostic-Diagnostic Systems, COA Animal Origin Information Position Statement", located on the Internet website address at <http://www.bd.com/regdocs/>.

John Gerlich  
Vice President,  
Quality Management and  
Regulatory Compliance  
Signature Date: 2011/02/17