

Model	Description	Measurement Range	Resolution	Calibration
ST10	Pen pH Meter; 0.1 pH	0.0 - 14.0	0.1 pH	0.1 pH
ST20	Pen pH Meter; 0.1 pH with temp display	0.00 - 14.00	0.01 pH	0.05 pH
ST10R	Pen ORP Meter; -1000 mV - 1000 mV	-1000mV - 1000mV	1 mV	2 mV
ST20R	Pen ORP Meter; -1000 mV - 1000 mV with temp display	-1000mV - 1000mV	1 mV	2 mV
ST10C-A	Pen conductivity Meter; 0.0 - 199.9 us/cm	0.0 - 199.9 us/cm	0.1 us/cm	±2.5% FS
ST10C-B	Pen conductivity Meter; 0.0 - 1999 us/cm	0.0 - 1999 us/cm	1 us/cm	±2.5% FS
ST10C-C	Pen conductivity Meter; 0.0 - 19.99 ms/cm	0.0 - 19.99 ms/cm	10 us/cm	±2.5% FS
ST20C-A	Pen conductivity Meter; 0.0 - 199.9 us/cm with temp display	0.0 - 199.9 us/cm	0.1 us/cm	±1.5% FS
ST20C-B	Pen conductivity Meter; 0.0 - 1999 us/cm with temp display	0.0 - 1999 us/cm	1 us/cm	±1.5% FS
ST20C-C	Pen conductivity Meter; 0.0 - 19.99 ms/cm with temp display	0.0 - 19.99 ms/cm	10 us/cm	±1.5% FS
ST10S	Pen Salinity Meter; 0.0 - 10.0 ppt	0.0 - 10.0 ppt	0.1 ppt	±2.5% FS
ST20S	Pen Salinity Meter; 0.0 - 10.0 ppt with temp display	0.0 - 80.0 ppt	0.1 ppt	±1.5% FS
ST10T-A	Pen TDS Meter; 0.0 - 100.0 mg/L	0.0 - 100.0 mg/L	0.1 mg/L	±2.5% FS
ST10T-B	Pen TDS Meter; 0.0 - 1000 mg/L	0.0 - 1000 mg/L	1 mg/L	±2.5% FS
ST20T-A	Pen TDS Meter; 0.0 - 100.0 mg/L with temp display	0.0 - 100.0 mg/L	0.1 mg/L	±1.5% FS
ST20T-B	Pen TDS Meter; 0.0 - 1000 mg/L with temp display	0.0 - 1000 mg/L	1 mg/L	±1.5% FS