



Quality Certificate

VWR® NITRILE EXAMINATION GLOVE

As per the manufacturer, the below product meets the following criteria:

North American Catalog No :	82026-428
Lot No :	PT234R14P004
Description :	VWR GLOVE NITRILE PF L PK100
Date of Manufacturing :	December 2014
Product Shelf Life :	November 2019
Regulatory :	FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

Accuracy:	Water tight test:	Inspection level = G1;	AQL 1.5
	Visual inspection:	Inspection level = G1;	AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
	Dimensional:	Inspection level = S2;	AQL 4.0
	Physical Property:	Inspection level = S2;	AQL 4.0
	Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
	Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
	Glove count:	Inspection level = 5 dispensers;	Tolerance 100 +/- 2 pieces

Material: Nitril BUTADIENE Rubber, 100 % Latex Free

Sterility: N/A

Non- Pyrogenic: N/A

RNase/ DNase Free: N/A

Cytotoxicity: N/A

BSE TSE Statement: N/A

Signed :



Ken Crossley
Manager
Quality Assurance

Date : 03 Dec 2014