

## Certificate of Analysis

**Product:** 4X SDS Sample Buffer  
**Novagen Material Number:** 70607-3  
**Kit Batch Number:** 2642642  
**Component:** 2 ml 4X SDS Sample Buffer

**Storage:** Store at -20°C

**Buffer Formulation:** 250 mM Tris-HCl pH 6.8, 6% SDS, 300 mM DTT, 30% glycerol, and 0.02% bromophenol blue

**Test Conditions:** 4X SDS Sample Buffer has been tested to yield the expected banding pattern on Coomassie blue stained gels when 5 µl of Perfect Protein Markers or 10 µg of crude bacterial lysate were loaded on SDS PAGE mini-gels. There were no extraneous bands when sample buffer alone was loaded on gels.

**Certified by:** B. Fischer, Quality Manager

**Date:** 15-Dec-2015

### For Research Use Only

**Warranty:** EMD Chemicals, Inc., Novagen brand warrants that the product will meet or exceed these specifications when used (under normal conditions) in your laboratory. We will promptly replace the product free of charge if the product does not conform to these specifications. Our obligation and your sole remedy is limited to such replacement in the event that the product proves to be defective. No other warranties are expressed or implied, including, but not limited to, the implied warranties of merchantability and fitness for a particular purpose. EMD Chemicals, Inc., Novagen brand is not liable for consequential damages.

---

Shipping temperature may differ from storage temperature. This does not alter the performance of the product.

EMD Millipore Corporation | 28820 Single Oak Dr., Temecula, CA 92590 USA

Technical Support NA +1-800-221-1975 | email: [www.millipore.com/techservices](http://www.millipore.com/techservices) | [www.novagen.com](http://www.novagen.com)

Technical Support All Other Countries - Contact Your Local Office

**FOR RESEARCH USE ONLY.** Not for use in diagnostic procedures. Not for human or animal consumption. Purchase of this product does not include any right to resell or transfer, either as a stand-alone product or as a component of another product. Any use of this Product for purposes other than research is strictly prohibited.