

# **Certificate of Quality**

## **VWR® LABORATORY BOTTLE**

As per the manufacturer, the below product meets the following criteria:

VWR Catalog Number	414004-115		
Lot Number	VHR130116		
Description	VWR BOTTLE HDPE WM 1000ML PK6		
Date of Manufacture:	January 13, 2016.		
Country of Origin: India.			

## **Quality System Compliance**

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2008 & ISO13485:2003 Quality Management System. Inspection records are reviewed and signed for product release.

### **QC Testing**

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

Test Name	Test Reference	Accuracy	Result
Leak Test	TPPL/TEST/003	General Insp. Level- I, AQL- 1.5	Leak proof
Autoclaving	TPPL/TEST/001	General Insp. Level- I, AQL- 1.5	NA
Dimensional	QSF-QC-02	General Insp. Level- I, AQL- 1.5	Within tolerance

#### Product Specifications

Material: Bottle-High Density Polyethylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.

> Polypropylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed. Closure-

**BSE/TSE Statement:** The above mentioned product is manufactured without the use of Raw materials of animal origin.

Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's

knowledge.

Not Applicable.

Not Applicable. Non-Pyrogenic: **DNase & RNase Free:** 

Cytotoxicity: Not Applicable.

**Latex Free Statement:** No materials containing latex or natural rubber are used in the manufacturing, handling and

packaging processes for this product.

Signed:

Ken Crossley Manager

Quality Assurance Date: January 13, 2016.