



ISO 9001 CERTIFIED
ISO 13485 CERTIFIED

Corporate Headquarters

AMRESCO, LLC

28600 Fountain Parkway
Solon, OH 44139 USA
TEL: 440-349-1199 FAX: 440-349-1182
WWW.AMRESCO-INC.COM
Email: info@amresco-inc.com

CERTIFICATE OF ANALYSIS

EZVISION Three DNA Dye & Buffer, 6X

Code: N313
Batch Number: 3136C342

Chemical Formula:	Not Applicable	Manufacture Date:	01/19/2016
Molecular Weight:	Not applicable	Reassay Date:	01/31/2018
CAS #:	Not Applicable		
Appearance:		Storage:	Cold
Dark purple, slightly viscous liquid.		Grade:	BIOTECHNOLOGY GRADE

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Abs.@ 525nm (1:500mL, Water)	0.4 - 0.8	0.5	PASS
Abs.@ 588-594nm (1:500mL, Water)	0.5 - 0.8	0.6	PASS
Abs.@ 635-641nm (1:500mL, Water)	0.55 - 0.9	0.64	PASS
Clarity	PASS / Fail	PASS	PASS
DNase	NONE / Fail	NONE	PASS
Electrophoresis (P/F)	PASS / FAIL	PASS	PASS
Protease	NONE / Fail	NONE	PASS

Spec Set: N313BIO

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

Internal ID #: 546

Signature:

Title: QC Manager

Date Printed: 06/16/2016

Page 1 of 1