



ISO 9001 CERTIFIED
ISO 13485 CERTIFIED

Corporate Headquarters

AMRESKO, LLC

28600 Fountain Parkway
Solon, OH 44139 USA
TEL: 440-349-1199 FAX: 440-349-1182
WWW.AMRESKO-INC.COM
Email: info@amresco-inc.com

CERTIFICATE OF ANALYSIS

Bradford Reagent

Code: E530
Batch Number: 0075C095

Chemical Formula:	Not Applicable	Manufacture Date:	10/02/2015
Molecular Weight:	Not applicable	Reassay Date:	10/31/2017
CAS #:	Not Applicable		
Appearance:		Storage:	Cold
Clear blue liquid.		Grade:	BIOTECHNOLOGY GRADE

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Abs.@ Peak A	0.95 - 1.6	1.51	PASS
Abs.@ Peak A/Abs.@ 595nm	>= 1.70	2.63	PASS
Bradford Protein Assay	PASS / Fail	PASS	PASS
Characteristic Pattern	PASS / Fail	PASS	PASS
Peak A	463 - 467 nm	466 nm	PASS
pH @25C	0.5 - 1.5	0.6	PASS

Spec Set: E530BIO

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

Internal ID #: 546

Signature:

Title: QC Manager

Date Printed: 07/06/2016

Page 1 of 1