


Date of Certificate: 04/05/2016
Product Number: 2019-2000
Product Lot Number: 1170802
Specified Dose Range: 20.0 - 45.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1699332	03/29/2016	280	CS	23.3	29.3
1697929	03/30/2016	280	CS	23.4	29.9
1699763	03/30/2016	595	CS	23.5	29.3
1699750	03/30/2016	420	CS	23.7	29.4
1700703	04/04/2016	175	CS	23.8	29.5



Barry Srolis
Sr. Quality Manager