

## **Certificate of Compliance**

**Corning Incorporated - Life Sciences** Antigua Carretera A Roma KM 3.6 San Nicolas De Los Garza NL 66480 MEX www.corning.com/lifesciences Refer to website for regional contact information.

Page: 1 / 1

6140 FNL 65MM Product Name

Manufacture Date : 2016-07-07 Catalog Number 6140-65

Lot ID 18916001

Unconfirmed Validity Expiration Date

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Glass Type - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

BSE/TSE - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass Packaging Inspection - Pass Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

Racelis Acosta Reynoso

Quality Manager

Creation Date: 2016-08-16 22:30:09