

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 7065-1
Lot Number: 22613049
Product Description: PYREX® Pipet, Reusable Glass, Measuring, 1 mL in 1/10 mL
Color-Coded, with Colored Markings

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:

Type I Class B - Alumino-Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:

Capacity - 1.00 mL ± 0.02 mL at any mark

TSE/BSE:

Materials of animal origin are not used in the manufacture of this product.

This product is designed to comply with the requirements of ASTM E1293 (Specification for Glass Measuring Pipets). This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:

Visual Attributes
Dimensional
Packaging
Capacity

Results:

Pass
Pass
Pass
Pass

Revision Date:

13 AUG 2013



Frigyes Szendro Kassa
Quality Manager

Rev No: 1