

Date of Certificate: 04/11/2017
Product Number: 5000-0020
Product Lot Number: 1198151
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1925481	03/27/2017	156	CS	20.7	26.9
1930083	04/04/2017	134	CS	19.4	26.3
1927128	04/04/2017	22	CS	20.6	26.4



Barry Srolis
Sr. Quality Manager