

Certificate of Compliance

Corning Incorporated - Life Sciences Antigua Carretera A Roma KM 3.6 San Nicolas De Los Garza NL 66480 MEX www.corning.com/lifesciences Refer to website for regional contact information.

Page: 1 / 1

9825 RGSCT 13X100 TUBE COMP Product Name W/ SCREW CAP

Catalog Number 9825-13 Manufacture Date : 2016-07-27

Lot ID 20916021

Unconfirmed Validity Expiration Date

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Glass Type - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass Packaging Inspection - Pass Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

Racelis Acosta Reynoso Quality Manager

Creation Date: 2016-10-13 22:30:19