

Corning Incorporated - Life Sciences
Avenida Industrial del Norte S/N
Parque Industrial del Norte
Reynosa TAMP 88736 MEX
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name	: Tube 14mL 17x100mm RBtm PP25pk	Manufacture Date	: 2017-11-20
Catalog Number	: 352059		
Lot ID	: 32417070		
Expiration Date	: 2022-11-30		

Quality Management System - Complies with the current version of the ISO 9001 Standard and the FDA CFR 21 Part 820, current Good Manufacturing Practices (cGMP).

BSE/TSE - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Non-Pyrogenic - Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing". The acceptance level for product is ≤ 0.10 EU/ml or ≤ 4 EU/device.

USP Class VI Testing - All material resin is tested, qualified and shown to be non-toxic as established in the Standards USP Class VI Chapter<87>, "Biological reactivity Tests, in Vitro" and Chapter<88>, "Biological Reactivity Tests, in vivo".

DNase/RNase Free - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is 10^{-7} Kunitz units/uL for DNase and 10^{-9} Kunitz units/uL for RNase.

Sterility - Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10^{-3} .

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release.

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



JOSE DOMINGUEZ
Quality Manager