

## Certificate Of Analysis

**Whatman™**

### Product Information

Grade: 1CHR

Product Number: 3001-917

Product Name: GR 1CHR 46x57CM 100/PK

Lot Number/Serial Number: 13598053

Date of Manufacture: 10-Oct-2017

### Quality Control

| Test                    | Range   | Results |
|-------------------------|---------|---------|
| Grammage,gsm            | 80-93   | 88      |
| Thickness, um           | 160-196 | 170     |
| Porosity, S/200ml/sq.in | 13-31   | 20      |

### Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to ISO9001:2008 and tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

### Electronic signature

This document has electronically produced and is valid without a signature.

Version AC



imagination at work

[www.gelifesciences.com](http://www.gelifesciences.com)  
**GE Healthcare UK Limited**  
Amersham Place Little Chalfont  
Buckinghamshire HP7 9NA UK