

Certificate of Analysis

Page: 1 of 3

Becton Dickinson and Company BD Diagnostic Systems PO Box 999 Sparks MD 21152-0999 US

Product Name : Bottle Mueller Hinton Agar 500G

Catalog Number : 225250 Manufacture Date: 2017/11/11

Batch Number : 7363834
Expiration Date : 2020/03/31

01. Dehydrated Medium Appearance: Beige, free-flowing, homogeneous

02. Solubility: 3.8% solution, soluble in distilled or deionized water on boiling

03. Prepared Plate Appearance: Clear to slightly opalescent

04. Certificate of Suitability issued by the European Directorate for the Quality of Medicines (EDQM): CEP 2000-317

05. Cultural Response: Medium was prepared per label instructions. Plates were inoculated and antibiotic disks were tested as recommended by current CLSI. The diameters of the zones around designated disks with the appropriate test culture fell within the CLSI recommended zone ranges.

TEST ORGANISMS ATCC® Enterococcus faecalis 33186 Escherichia coli 25922 Pseudomonas aeruginosa 27853 Staphylococcus aureus 25923

Characteristic	Unit	Value	Lower Limit	Upper Limit
pH at 25°C : Bulk Lot Number :	_	7.3 7285906	7.2	7.4

Animal source Country of Origin		Tissue Category BIC SIC ABC			
				ABC	
Avian	China	IV	IV	NDF	
Bovine	Australia	IV	IV	IB	
Bovine	New Zealand	IV	IV	MLK	
Bovine	USA	III	III	IB	
Bovine	USA	III	IV	IC	
Bovine	USA	IV	IV	IB	
Bovine	USA	IV	IV	NDF	
Porcine	USA	III	III	IB	

For an avian origin ingredient used in the manufacture of this product, the BD supplier is not able to confirm animal health. The supplier has confirmed that the ingredient undergoes extensive processing that includes exposure to strong acid, filtration, crystallization, and drying.

The Batch Number on this certificate is synonymous with the Lot Number

Creation Date: 2018/01/30 16:30:45



Certificate of Analysis

Becton Dickinson and Company BD Diagnostic Systems PO Box 999

Sparks MD 21152-0999 US Page: 2 of 3

Product Name : Bottle Mueller Hinton Agar 500G

Catalog Number : 225250 Manufacture Date: 2017/11/11

Batch Number : 7363834
Expiration Date : 2020/03/31

shown on the product label.

BD Diagnostics - Diagnostic Systems products are manufactured in ISO 9001:2008 Registered facilities. In addition, BD Diagnostics - Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), are regulated by the FDA's Quality System Regulations (QSRs), and are also ISO 13485:2003 Registered. This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at http://www.bd.com/regdocs. For each batch of finished product that contains animal origin raw materials, the COA shows the animal origin data from the individual lots of animal origin raw materials used, as provided by the raw material suppliers.

At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at http://www.bd.com/regdocs under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program.

For complete details on animal origin information, refer to "BD Position Statement - BD Diagnostic-Diagnostic Systems, COA Animal Origin Information Position Statement", at http://www.bd.com/regdocs under "Position Statements".

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see www.bd.com/ds/technicalCenter/regulatory.asp.

Creation Date: 2018/01/30 16:30:45



Certificate of Analysis

Page: 3 of 3

Becton Dickinson and Company BD Diagnostic Systems PO Box 999

Sparks MD 21152-0999 US

Product Name : Bottle Mueller Hinton Agar 500G

Catalog Number : 225250 Manufacture Date: 2017/11/11

Batch Number : 7363834 Expiration Date : 2020/03/31

Charlette Danserflyen

Charlotte Dannenfelser
BD Life Sciences - Diagnostic Systems
Quality Director, Microbiology

Signature Date: 2018/01/30

Creation Date: 2018/01/30 16:30:45