CORNING

Certificate of Compliance

Corning Incorporated - Life Sciences
Antigua Carretera A Roma KM 3.6
San Nicolas De Los Garza NL 66480 MEX
www.corning.com/lifesciences
Refer to website for regional contact information.

Product Name	: 7100 PIPET 10ML		
Catalog Number	: 7100-10	**Manufacture Date**	: 2017-08-28
Lot ID	: 24017035		
Expiration Date	: Unconfirmed Validity		

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Glass Type - Alumino-Borosilicate, Type I, Class B, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

Volumetric Tolerance - Volumetric Accuracy at fill line +/- 0.02 mL.

Product Specification Standards - Product made in accordance with:
ASTM E542 - Specification for Calibration of Laboratory Volumetric Apparatus
ASTM E694 - Standard Specification for Laboratory Glass Volumetric Apparatus, Class A Criteria
ASTM E969 - Standard Specification for Glass Volumetric (Transfer) Pipets

BSE/TSE - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:
Visual Inspection - Pass
Packaging Inspection - Pass
Dimensional Test - Pass
Capacity Verification - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

Miguel Flores
Quality Manager

Creation Date: 2017-12-09 16:38:09