



Quality Certificate

VWR® ERGONOMIC HIGH-PERFORMANCE SINGLE-CHANNEL VARIABLE VOLUME PIPETTERS

As per the manufacturer, this product meets the following criteria:

North American Catalog No:	89079-974
European Catalog No:	613-5265
Serial Number:	742764762
Description:	VWR PIPETTE EHP VAR 100-1000UL

The enclosed liquid handling instrument has been manufactured in a facility that employs a registered and certified quality management system. This system ensures that the development and production of this product has been strictly controlled and that it meets the highest quality standards. Each instrument passes stringent validation and Control Procedures in accordance with EN ISO 8655.

Volume (µl):	100	1000
Accuracy (%):	0.8415	0.4695
Precision (%):	0.1009	0.0173
Status:	Passed	Passed

Calibrated and tested using the gravimetric method with distilled water:

Tested by: C11

Condition of measurements:

Basis of adjustment: Ex; Reference temperature e: 20°C; Relative air humidity: 50%; Barometric pressure: 101kPa

This document has been produced electronically and is valid without a signature.

Signed on behalf of VWR International:

Michael J. Mascali, RPh, FASCP, ASQ CMQ – OE
SVP Compliance
VWR International LLC

Date: October 16, 2017