

**Date:** June 27, 2018  
**Catalog Number:** 89130-898  
**Lot Number:** 16318030  
**Product Description:** PIPETTE,VWR,10ML,PPW,PS,S,IND,  
**Manufacturing Date:** 6/12/2018 **Expiration Date:** 6/11/2021

**Component Materials:**

Pipette - Polystyrene, meets USP Cytotoxicity Test <87>.  
Mouthpiece Plug - Polyester, meets USP Cytotoxicity Test <87>.  
Ink - meets USP Cytotoxicity Test <87>.

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product is deemed animal free by virtue of not containing materials of animal origin and/or complies with the latest revision of EMA/410/01 section 6.4.

**Sterilization:**

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 "Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization". Sterility Assurance Level: SAL  $10^{-3}$

**Pyrogens:**

The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: "Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing". Results:  $\leq 0.1$  EU/mL ( $\leq 4$  EU/device)

**DNase/RNase Free:**

The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: "Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing". Results:  $\leq 0.1$  EU/mL ( $\leq 4$  EU/device)

**Volumetric Accuracy:**

This product has been tested and is accurate within +/- 2% at stated full volume. This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Volumetric Accuracy	Pass
Packaging	Pass



Jason Jens  
Plant Quality Manager  
Rev. No. 6